OpsAuditHub opsauditihub.com	DOC ID: **SAF-0171**	REVISION: **Rev 2.4 (2026)**
	STANDARD: **The Joint Commission (TJC) EC.02.01.01 & AHA Emergency Cardiovascular Care Guidelines**	AUDIT FREQUENCY: **Daily Pre-Shift**

Crash Cart Daily Inspection Daily Log Sheet

Operational Scope: Mandatory daily verification of emergency crash cart readiness to ensure immediate clinical accessibility during cardiac or respiratory arrest. Mitigates patient morbidity risks through strict compliance with life-safety inventory and tamper-evident security protocols.

MANDATORY PPE:	Nitrile Medical Examination Gloves (ASTM D6319), Fluid-Resistant Procedural Mask, Safety Glasses with Side Shields (ANSI Z87.1)
PREREQUISITES:	Verify cart is in its designated, unobstructed primary staging location. | Ensure clinical staff present during the inspection are properly trained in emergency cart deployment.

EXTERIOR SECURITY & MECHANICAL INTEGRITY

#	COMPONENT / ITEM	VERIFICATION CRITERIA & ACCEPTANCE STANDARD	SEVERITY	STATUS
1	**Tamper-Evident Lock / Seal**	Lock is intact, unbroken, and the serial number matches the daily log sheet record exactly.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
2	**Cart Exterior & Wheels**	Cart exterior is clean, drawers slide smoothly, and all four swivel casters lock and roll without binding.	**Warning**	[ ] PASS [ ] FAIL [ ] N/A
3	**Oxygen Cylinder Supply**	Tank is securely mounted to the cart, pressure gauge reads at or above 1,000 PSI, and a wrench is attached.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
4	**Suction Unit Functionality**	Suction machine powers on, battery charge indicator shows full, and canister tubing is unobstructed.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A

AIRWAY & RESUSCITATION EQUIPMENT

#	COMPONENT / ITEM	VERIFICATION CRITERIA & ACCEPTANCE STANDARD	SEVERITY	STATUS
5	**Bag-Valve-Mask (BVM) Resuscitators**	Includes adult and pediatric sized BVMs, unopened in packaging, fully functional with clear masks.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
6	**Airway Management Supplies**	Laryngoscope handle with working batteries, assorted blades (Miller/Macintosh), and endotracheal tubes are present and unexpired.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
7	**Suction Catheters & Yankauer**	Sterile Yankauer tips and suction catheters of varying French sizes are present and within expiration dates.	**Warning**	[ ] PASS [ ] FAIL [ ] N/A

MONITORING, DEFIBRILLATION & PHARMACOLOGY

#	COMPONENT / ITEM	VERIFICATION CRITERIA & ACCEPTANCE STANDARD	SEVERITY	STATUS
8	**Defibrillator / Monitor Unit**	Unit passes daily self-test, battery indicator shows fully charged, and adult/pediatric pads are unexpired.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
9	**Emergency Medications Tray**	All ACLS/PALS emergency medications (Epinephrine, Amiodarone, etc.) are present, fully stocked, and none are expired.	**Critical (Immediate Out of Service)**	[ ] PASS [ ] FAIL [ ] N/A
10	**IV Access & Fluids Supplies**	IV catheters, extension sets, normal saline flush syringes, and administration sets are present and unexpired.	**Warning**	[ ] PASS [ ] FAIL [ ] N/A

Certified Clinical Inspector / RN Signature: _____	**Date / Time:** _____	**Badge / Cert #:** _____
Clinical Nurse Manager / Safety Officer Sign-off: _____	**Review Date:** _____	**Audit Status:** [ ] Approved [ ] Rectify

COMPLIANCE NOTICE: This document must be archived for a minimum of 12 months in accordance with healthcare regulatory and workplace safety recordkeeping guidelines.

DIGITAL INSPECTION UPGRADE: Run this audit on mobile/tablet, attach photos & automate compliance records. **Start Free Mobile Audit →**